

• The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.

• Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).

• Standards are certifed (+/-) 0.5% of the stated value, unless otherwise stated.

• All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.

• Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result,"

NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).





Run 33, "P70274 L082922 [1000µg/mL in MeOH]"

Run Length: 40.00 min, 23999 points at 10 points/second. Created: Sat, Sep 3, 2022 at 4:50:17 PM. Sampled: Sequence "090222-GC4M2", Method "GC4-M2". Analyzed using Method "GC4-M2".

Comments

GC4-M2 Analysis by Candice Warren Column ID SPB-5 L#60062-01A 30 meter x 0.53mm x 1.5um Film Thickness. Flow rates; Total Flow = 300 ml/min, Helium (carrier) = 6.5 mL, Helium (make-up) = 25 mL. Hydrogen (detector) = 30 mL, Air (detector) =360 mLOven Temp 1 = 50°C (1 min). Rate = 10°C/min, Oven Temp 2 = 300°C (14 min), Total Run Time = 40 Minutes. Injector Temp = 250°C. FID Temp = 300°C, FID Signal = eDaq Channel 1. Gas Chromatograph = HP 5890, Auto Sampler = HP 7673, Standard Injection = 0.5 uL, Range =4

