

• The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.

• Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).

• Standards are certifed (+/-) 0.5% of the stated value, unless otherwise stated.

• All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.

• Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result,"

NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).





Run 39, "P79181 L042419 [1000µg/mL in Acetone]"

Run Length: 35.00 min, 20999 points at 10 points/second. Created: Thu, Apr 25, 2019 at 4:44:48 PM. Sampled: Sequence "042419-GC9M1", Method "GC9-M1". Analyzed using Method "GC9-M1 [2]".

Comments

GC9-M1 Analysis by Candice Warren Column ID Rtx-5 30 meter x 0.53mm x .5um Film Thickness Flow rates; Total Flow = 300 ml/min, Helium (carrier) = 6.5 mL, Helium (make-up) = 25 mL, Hydrogen (detector) = 30 mL, Air (detector) =360 mL Oven Temp 1 = 50°C (1 min), Rate = 10°C/min, Oven Temp 2 = 300°C (9 min), Total Run Time = 35 Minutes. Injector Temp = 200°C, FID Temp = 300°C, FID Signal = eDaq Channel 1. Gas Chromatograph = HP 5890, Auto Sampler = HP 7673, Standard Injection = 0.5 uL, Range = 4

