



nsi lab solutions

protect more. test more.

7212 ACC Blvd. | Raleigh, NC 27617 | 800.234.7837

# Certificate of Analysis

## 4,4'-DDD

**Catalog Number:** 92

**Lot Number:** 020717

**Manufacture Date:** 02/07/2017

**Certified Date:** 02/13/2017

**Expiration:** 02/28/2027

**Solvent:** Methanol

**Hazards:** Irritant, Flammable,  
Carcinogen

| <u>Analyte</u> | <u>CAS</u> | <u>Analyte Purity</u> | <u>Gravimetric Concentration (ug/mL)</u> |
|----------------|------------|-----------------------|--|
| 4,4'-DDD       | 72-54-8    | 99.0%                 | 1000.5 ± 1.75                            |

### Packaging, Storage, Instructions For Use

This reference standard is packaged in a flame-sealed ampule and **must be stored at 2 to 8°C**. To use this standard, mix it gently by inversion. Inspect for precipitate. If present, sonicate for a few minutes to redissolve. Open the ampule and withdraw an aliquot appropriate for your application.

### Traceability Information

**Analyte Source Materials:** The highest purity analyte source materials are used in the manufacture of this standard. The actual purity is referenced above.

**Method:** This standard was verified gravimetrically.

**Balance:** All analytical balances are calibrated on a semiannual basis by an ISO 17025 accredited calibration laboratory and are traceable to NIST. Traceable Calibration Certificate available upon request.

All balances are checked daily by an in-house standard operating procedure. The weights used for this daily verification are calibrated annually by an ISO 17025 accredited calibration laboratory and are certified traceable to NIST. Certificate of Calibration and Traceability available upon request.

**Thermometer:** All thermometers are NIST traceable through thermometers that are calibrated annually by an ISO 17025 accredited calibration laboratory.

**Glassware:** All glassware used in the manufacture of our standards is Class A. An in-house standard operating procedure is used to verify all glassware prior to it being placed into service. Volumetric pipetors are calibrated every four months by an ISO 17025 accredited calibration laboratory.



ISO Guide 34:2009  
Certificate AR-1571

ISO 9001:2008 UL Registered Firm - Certificate # 10002343 QM08



ISO/IEC 17025:2005  
Certificate AT-1690

**Intended Uses**

- Calibration of analytical instruments
- Validation of analytical methods
- Preparation of working level reference materials, i.e. "check standards"
- Detection limit studies

**Catalog Number: 92**

**Lot Number: 020717**

**Homogeneity**

This standard was thoroughly mixed in production and is guaranteed homogenous.

*Kenneth Grzybowski*

---

Kenneth Grzybowski, Technical Organics Manager

*Mark Hammersla*

---

Mark Hammersla, President