

分析证明书

色度 质控样



产品编号: CQC-087

批号: D0421

保质期: 2023-04-30

基质/溶剂: 水

危害: 刺激

项目	能力验证 统计值 CU	确认值 CU	接受区间 CU
色度	12.8	14.0 ± 0.1302	7.34 - 20.7

该质控样品在 由美国NSI公司配制, 配制过程符合 ISO9001, ISO17025及 ISO17034 认证
接受区间根据当前美国能力验证行业标准设置
能力验证统计值 为 剔除不合格回报值后的实验室内能力验证结果统计值
该样品可用于: 方法验证、检测极限研究、能力验证等

储存及使用说明

2-8℃保存

- 1, 即用型包装, 直接使用, 不需要稀释
- 2, 该样品适用于所有色度分析方法
- 3, 使用前翻转几次以混匀
- 4, 根据方法汲取适量使用
- 5, 每次用完后旋紧瓶盖

溯源

原料: 分析项目原料选用可用的最高纯度原料用于配制该样品。如有相应的 NIST标准物质可用, 原料纯度及不确定度会与其对照分析校验

天平: 所有天平按ISO17025校准实验室认证要求每季度校准一次, 溯源至NIST。所有天平每天按照内部标准操作程序查验, 查验所用砝码按17025认证要求每年校准一次。

温度计: 所有温度计溯源至 NIST, 每年校准一次

玻璃器皿: 此样品配制过程中涉及的所有玻璃器皿为 A 级。所有玻璃器皿启用前经过内部标准操作程序校验。移液器按17025认证要求每月校准一次。

均匀性/稳定性/保质期

该标物生产过程中已充分混匀, 批次均匀性按要求随机取样分析建立。该标物稳定性基于短期及长期对确认浓度的监测结果。保质期基于长期监测结果确保保质期内有效

不确定度

不确定度为 95%置信区间扩展系数 K=2.

ISO 9001:2015 UL Registered Firm – Certificate # 10002343 QM15



ISO 17034:2016 - Certificate AR-1571



ISO/IEC 17043:2010 - Certificate AP-1693



ISO/IEC 17025:2017 - Certificate AT-169C

Certificate of Analysis

Color QC CRM

Catalog Number: CQC-087

Lot Number: D0421

Expiration Date: 2023-04-30

Matrix: WATER

Hazards: Irritant

<u>Analyte</u>	<u>Study Mean</u> CU	<u>Certified Concentration</u> CU	<u>Acceptance Limits</u> CU
Color	12.8	14.0 ± 0.1302	7.34 - 20.7

This quality control sample was manufactured by NSI Lab Solutions following quality procedures meeting the requirements of ISO 9001, ISO 17025, and ISO 17034. Acceptance limits are set at current NELAC standards. The study mean is set at the mean of an interlaboratory proficiency testing study with outlier rejection. This sample is intended to be used to validate analytical methods, for detection limit studies, and analyst proficiency testing.

Storage Instructions For Use

Store tightly capped at room temperature (15-30°C).

This sample is a whole volume sample. No dilutions are required.

This sample is applicable to all approved color analysis methods.

Mix the sample by inversion prior to opening.

Measure an aliquot appropriate to your method.

Cap the sample tightly after each use.

This is your sample for analysis.

Traceability Informatio

Analyte Source Materials: The highest purity analyte source materials are used in the manufacture of this sample. Analyte source material purity and associated uncertainty has been analytically verified against appropriate NIST SRMs,

Balance: All analytical balances are calibrated on a semiannual basis by an ISO 17025 accredited calibration laboratory and are traceable to NIST. Traceable Calibration Certificate available upon request.

All balances are checked daily by an in-house standard operating procedure. The weights used for this daily verification are calibrated annually by an ISO 17025 accredited calibration laboratory and are certified traceable to NIST. Certificate of Calibration and Traceability available upon request.

Thermometer: All thermometers are NIST traceable through thermometers that are calibrated annually by an ISO 17025 accredited calibration laboratory.

Glassware: All glassware used in the manufacture of our samples is Class A. An in-house standard operating procedure is used to verify all glassware prior to it being placed into service. Volumetric pipetors are calibrated every four months by an ISO 17025 accredited calibration laboratory.

ISO 9001:2015 UL Registered Firm – Certificate # 10002343 QM15





Catalog Number: CQC-087

Lot Number: D0421

Homogeneity/Stability/Expiration

This quality control sample was thoroughly mixed in production. Batch homogeneity was established through analyses of samples chosen at random. The stability of this quality control sample is based on short-term and long-term monitoring of the certified concentration. The expiration date is guaranteed to be valid from the manufacture date and is based on results of long-term monitoring.

Uncertainty

The \pm uncertainty associated with the certified concentration is the expanded uncertainty at 95% confidence interval (CI) with K=2. This expanded uncertainty incorporates contributions from manufacturing, homogeneity, and stability.

Ewart Morris

Ewart Morris, Inorganics Technical Manager

Mark Hammersla

Mark Hammersla, President

ISO 9001:2015 UL Registered Firm – Certificate # 10002343 QM15

