

Certificate of Analysis

Sulfide in Soil

Catalog Number: SQCI-018
Lot Number: S0222
Manufactured Date: 01/31/22
Certified Date: 01/31/22

Expiration: 01/31/2024
Matrix: Dilute NaOH
Hazards: Irritant

<u>Analyte</u>	<u>Study Mean</u> (mg/kg)	<u>Certified Concentration</u> (mg/kg)	<u>Acceptance Limits</u> (mg/kg)
Sulfide	315	423 ± 3.94	119 - 510

This QC sample contains acid-soluble Sulfide.

This quality control CRM was manufactured by NSI Lab Solutions following quality procedures meeting the requirements of ISO 9001, ISO 17025, and ISO 17034. Acceptance limits are set at current NELAC standards. The certified concentration is the gravimetric true value determined during manufacture, masses traceable to NIST. The study mean is set at the robust mean of an interlaboratory proficiency testing study with outlier rejection. This CRM is intended to be used to validate analytical methods, for detection limit studies, and analyst proficiency testing.

Storage & Instructions For Use

Required storage condition is 2-8°C. Keep away from light.

Weigh 25 g of the blank soil and place in extraction vessel.

Open ampule and add exactly 1 mL of solution directly to the blank soil.

The sample is now ready for immediate analysis.

Use an appropriate extraction and analytical method for Sulfide, assuming a 25 g sample will contain sufficient Sulfide for determination.

QC sample has been buffered to stabilize Sulfide. Some methods may require a pH adjustment to 7 prior to analysis.

Report results as mg/kg based on 25 g of soil. Do not correct for matrix spike recovery bias.

Traceability Information

Analyte Source Materials: The highest purity analyte source materials are used in the manufacture of this CRM. Analyte source material purity and associated uncertainty has been analytically verified against appropriate NIST SRMs,

Balance: All analytical balances are calibrated on a semiannual basis by an ISO 17025 accredited calibration laboratory and are traceable to NIST. Traceable Calibration Certificate available upon request.

All balances are checked daily by an in-house standard operating procedure. The weights used for this daily verification are calibrated annually by an ISO 17025 accredited calibration laboratory and are certified traceable to NIST. Certificate of Calibration and Traceability available upon request.

Thermometer: All thermometers are NIST traceable through thermometers that are calibrated annually by an ISO 17025 accredited calibration laboratory.





Catalog Number: SQCI-018

Lot Number: S0222

Glassware: All glassware used in the manufacture of our samples is Class A. An in-house standard operating procedure is used to verify all glassware prior to it being placed into service. Volumetric pipetors are calibrated every four months by an ISO 17025 accredited calibration laboratory.

Homogeneity/Stability/Expiration

This CRM was thoroughly mixed in production. Batch homogeneity was established through analyses of samples chosen at random. The stability of this CRM is based on short-term and long-term monitoring of the certified concentration. The expiration date is guaranteed to be valid from the manufacture date and is based on results of long-term monitoring.

Uncertainty

The \pm uncertainty associated with the certified concentration is the expanded uncertainty at 95% confidence interval (CI) with K=2. This expanded uncertainty incorporates contributions from manufacturing, homogeneity, and stability.

Ewart Morris

Ewart Morris, Inorganics Technical Manager